



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/531,912

05/08/2006

Wei Hu

37610-523N01

3778

23980

7590

01/27/2009

MINTZ, LEVIN, COHN, FERRIS, GLOVSKY AND POPEO, P.C

5 Palo Alto Square - 6th Floor

3000 El Camino Real

PALO ALTO, CA 94306-2155

EXAMINER

NGUYEN, BAO THUY L

ART UNIT

PAPER NUMBER

1641

MAIL DATE

DELIVERY MODE

01/27/2009

PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

## Office Action Summary

**Application No.**

10/531,912

**Applicant(s)**

HU, WEI

**Examiner**

Bao-Thuy L. Nguyen

**Art Unit**

1641

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 18 November 2008.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1, 2 and 6-47 is/are pending in the application.
- 4a) Of the above claim(s) 6-40 is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1, 2 and 41-47 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO/S5108)  
Paper No(s)/Mail Date 11/18/08
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

### **DETAILED ACTION**

#### ***Continued Examination Under 37 CFR 1.114***

1. A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(c), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(c) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on 11/18/2008 has been entered.

#### ***Status of the Claims***

2. Claims 3-5 have been canceled.
3. Claims 6-20 have been withdrawn.
4. Claims 1-2, 41-47 are pending.

#### ***Claim Rejections - 35 USC § 112***

5. The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

6. Claims 1, 2 and 41-47 are rejected under 35 U.S.C. 112, first paragraph, because the specification, while being enabling for a device where the sample deposition system comprises a sample deposition channel in a housing, does not reasonably provide enablement for a device comprising a carrier formed of a permeable material, one or more reagents disposed on the carrier and a sample deposition system operably associated with said carrier where said sample

deposition system delivers sample onto the carrier as a sample band of a first width, said sample band being essentially linear and the width being generally transverse to the path of sample flow.

The specification does not enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the invention commensurate in scope with these claims.

The specification as page 5 discloses a device comprising a housing and a carrier material. The housing is provided with a sample inlet that communicates with the carrier by way of a sample deposition channel. The sample deposition channel is adapted to deposit a sample as a generally linear band having its longer axis generally transverse to the path of sample flow on the carrier. The specification also teaches a device where the device has a sample receiving port that communicates with the carrier via a sample deposition means, which comprises a conduit formed within the upper member of the housing. The conduit feeds into sample reservoir which communicate with and feeds into a sample deposition channel from which sample is ultimately deposited onto the carrier. See page 14, lines 25-30.

No where in the specification is there a disclosure of a device where the sample deposition system is not part of a housing assembly.

***Claim Rejections - 35 USC § 112***

7. The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 1, 2 and 41-47 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

Claim 1, "transferase" should be --tranverse--.

Claim 2, "A device" should be changed to --The device-- for clarity.

### ***Double Patenting***

8. The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

9. Claims 1, 2 and 41-47 rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claims 1-4 and 8-12 of U.S. Patent No. 7,256,053.

Although the conflicting claims are not identical, they are not patentably distinct from each other because '053 discloses a device comprising a carrier for conducting the flow of sample, the carrier comprising a sample receiving zone and an analyte detection zone. The device further

comprises sample deposition system in flow communication with the sample receiving system and the carrier where the sample deposition system delivers sample to the carrier as a sample band that is essentially linear and generally transverse to the path of sample flow.

### ***Conclusion***

10. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Bao-Thuy L. Nguyen whose telephone number is (571) 272-0824. The examiner can normally be reached on Monday -- Thursday from 9:00 a.m. - 3:00 p.m.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Mark Shibuya can be reached on (571) 272-0806. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Application/Control Number: 10/531,912  
Art Unit: 1641

Page 6